



Viscometer 10-2,000,000 CP Model RV2M/2T




Model	Test Range	Spindles
RV-2M/2T	10-2,000,000 mpa.s	No. 1, 2, 3, 4

Specification

Model	RV-1	RV-1S	RV-1T	RV-2M	RV-2T
Measurement Range	10-100,000 <u>mpa.s/cp</u>			10-2,000,000 <u>mpa.s/cp</u>	
Display	LCD				
Spindles Specification	No. 1, 2, 3, 4 (No. 0 spindles can test 0.1-100 <u>mpa.s</u>)				
Spindles	6 / 12 / 30 / 60 rpm/min.			0.3 / 0.6 / 1.5 / 3 / 6 / 12 / 30 / 60 rpm/min.	
Interface	No	No	1. RS232 Interface And Edit Software 2. Temperature probe (0-100 degree)	No	1. RS232 Interface And Edit Software 2. Temperature probe (0-100 degree)
Automatic Switch	free selection of proper rotating speed or spindles number.				
Stable Cursor	Strip-type cursor fully charged displayed when reading basically stable				
Tolerance	<= ± 5%	<= ± 2%		<= ± 2%	
<u>Repeatability</u>	<= ± 2%	<= ± 1%		<= ± 1%	
Power	AC 100-240V/50-60 HZ				
Working Condition	Temperature 0-40 degree Relative humidity no more than 80%				
Attachments	1. No. 1,2,3,4 Spindles 2. No. 0 spindle is optional, if you need, pls contact with me 3. Adapter 4. Foundation Plate 5. Up-Down Frame 6. English Manual				
Outline Dimension (mm)	370×325×280 mm				
Packing (mm)	400×360×310 mm				
Net / Gross Weight (kg)	6.8 kg / 9kg				
H.S. Code	9027				





Certificates

CERTIFICATE OF CONFORMITY

Certificate No. : BKC-170502096C

Certificate of Conformity

Applicant : W&J INSTRUMENT CO., LTD.
No. 10 Qianjia Industrial Park, Changzhou City, China

Manufacturer : W&J INSTRUMENT CO., LTD.
No. 10 Qianjia Industrial Park, Changzhou City, China

Product : Viscometer

M/N : RM, RV, RM-1, RM-2, RV-1, RV-1S, RV-2M, RV-2T, RV-6M, RV-6T, RVDV.

Test Standard : EN 55014-1:2006+A1:2009+A2:2011
EN 61000-3-2:2014
EN 61000-3-3:2013
EN 55014-2:2015

The certificate of conformity is based on an evaluation of a sample of the above mentioned products. It does not imply an assessment of the whole production. It is possible to use CE marking to demonstrate the compliance with this EMC Directive 2014/30/EU. It is only valid in connection with the test report number: BKC-170502096R.



Shenzhen BKC Testing Co., Ltd.

6/F, Building 3, Zhouteng Industrial Park, Nanwan Street, Longgang District, Shenzhen, Guangdong, China. Tel: 4000-875-382 0755-32925341
E-mail: bkc@bkc-lab.com Certificate Search: <http://www.bkc-lab.com>

CERTIFICATE OF CONFORMITY

Certificate No. : BKC-170502096C

Certificate of Conformity

Applicant : W&J INSTRUMENT CO., LTD.
No. 10 Qianjia Industrial Park, Changzhou City, China

Manufacturer : W&J INSTRUMENT CO., LTD.
No. 10 Qianjia Industrial Park, Changzhou City, China

Product : Viscometer

M/N : RM, RV, RM-1, RM-2, RV-1, RV-1S, RV-2M, RV-2T, RV-6M, RV-6T, RVDV.

Test Standard : EN 60335-1:2012+A11:2014+AC:2014

The certificate of conformity is based on an evaluation of a sample of the above mentioned products. It does not imply an assessment of the whole production. It is possible to use CE marking to demonstrate the compliance with this LVD Directive 2014/35/EU. It is only valid in connection with the test report number: BKC-170502096R.



Shenzhen BKC Testing Co., Ltd.

6/F, Building 3, Zhouteng Industrial Park, Nanwan Street, Longgang District, Shenzhen, Guangdong, China. Tel: 4000-875-382 0755-32925341
E-mail: bkc@bkc-lab.com Certificate Search: <http://www.bkc-lab.com>