

UniBloc Moisture Analyzer MOC 63u

Easy-start mode:
fast response time

Larger pan size:
Ø95mm

Halogen heater:
quick and accurate
measurement

WindowsDirect -
easily integrate weighing results with
laboratory software without a special application

Built-in USB port

Compact design:
W202×D336×H157mm

**Easy operation using a
simple keypad**



Excellent performance for a wide variety of applications in multiple industries



Food

- Quality Assurance
- Harvest Inspection



Environmental

- Polluted Sludge Measurement
- Biofuel Measurement



Chemical

- Paint Quality Control
- Material Inspection



Pharmaceutical

- Drug Quality Assurance
- Cosmetics Inspection



■ MOC63u Specification

Capacity	Max	60 g
	Min	0.02g
Minimum readability		0.001g
		0.01/0.1% (Selectable)
Repeatability		0.15% (2g)
		0.05% (5g)
		0.02% (10g)
Drying Heater		Straight type halogen heater
Power		400W
Temperature range setting		50-200°C (1°C increments) (There is a time restriction when exceeding 180°C.)
Display		LCD with backlight
Pan size		Φ95mm
Dimension (W×D×H) mm		202 × 336 × 157
Weight		4kg
Operational temperature and humidity range		5 to 40°C, 85%RH or lower

Measurement modes	Standard (Easy start/Automatic end/Timed end)
	Rapid drying (Easy start/Automatic end/Timed end)
	Slow drying (Easy start/Automatic end/Timed end)
	Step drying (Easy start/Automatic end/Timed end)
Timer setting	1-120 minutes or continuous (max 12 hours)
Interface	RS-232C (9-pin connector) I/O port
	USB port
Measurement conditions data memory	10
Data memory	100
Temperature calibration kit	Option

■ Optional accessories

1	Printer EP-80
2	Printer EP-90
3	In-use protection cover for display (5 pieces)
4	Aluminum sheet
5	Fiberglass sheet
6	Temperature calibration kit
7	Sample pan (SUS)
8	RS-232C cable
9	USB connection cable
10	Halogen heater for replacement



Safety Precautions

Read instruction manual and understand before use of this instrument.

- Use this instrument for measurements in which water vaporizes from the sample under heating.
- The temperature of the heater installed in this instrument becomes higher than the set heating temperature for the sample.
- Any sample that is explosive, inflammable or may cause hazardous reaction under heating must not be measured with this instrument.



SHIMADZU CORPORATION. International Marketing Division

3. Kanda-Nishikicho 1-chome, Chiyoda-ku, Tokyo 101-8448, Japan Phone: 81(3)3219-5641 Fax: 81(3)3219-5710
<http://www.shimadzu.com/>

SHIMADZU SCIENTIFIC INSTRUMENTS, INC.
 7102, Riverwood Drive, Columbia, Maryland
 21046, U.S.A.
 Phone: 1(410)381-1227
 Fax: 1(410)381-1222
<http://www.ssi.shimadzu.com/>

SHIMADZU EUROPA GmbH
 Albert-Hahn-Strasse 6-10, 47269 Duisburg,
 F.R. Germany
 Phone: 49(203)7687-0
 Fax: 49(203)7666-25
<http://www.shimadzu.eu/>

SHIMADZU (ASIA PACIFIC) PTE LTD.
 79 Science Park Drive, #02-01/08, Cintech IV,
 Singapore Science Park 1, Singapore 118264
 Phone: (65)6778-6280
 Fax: (65)6779-2935
<http://www.shimadzu.com.sg/>

SHIMADZU INTERNATIONAL TRADING
 (SHANGHAI) CO., LTD.
 PUXI BRANCH
 Block E, No. 570 West Huaihai Road, Shanghai
 200052 P.R. China
 Phone: 86(21)2201-3888
 Fax: 86(21)2201-3666
<http://www.shimadzu.com.cn/>

SHIMADZU DO BRASIL COMERCIO LTDA.
 Avenida Marques de Sao Vicente,
 1771 Barra Funda
 CEP: 01139-003 - Sao Paulo, SP - Brazil
 Phone: 55(11)2134-1688
 Fax: 55(11)3611-2209
<http://www.shimadzu.com.br/>

SHIMADZU MIDDLE EAST & AFRICA FZE
 UAE HEAD OFFICE
 Warehouse No. RA08UC02, Jebel Ali, Dubai
 P.O. BOX 262081 United Arab Emirates
 Phone: 971(4)883-6668
 Fax: 971(4)883-6808

SHIMADZU SCIENTIFIC INSTRUMENTS
 (OCEANIA) PTY. LTD.
 Unit F, 10-16 South Street, Rydalmere N.S.W.
 2116, Australia
 Phone: 61(2)9684-4200
 Fax: 61(2)9684-4055
<http://www.shimadzu.com.au/>